

Notice of Allowability

Application No.

10/071,875

Examiner

George Manuel

Applicant(s)

AURICCHIO ET AL.

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to the information disclosure statement, filed 12/27/04.
2. ☒ The allowed claim(s) is/are 1-24.
3. ☒ The drawings filed on 08 February 2003 are accepted by the Examiner.
4. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).


* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO-1449 or PTO/SB/08),
Paper No./Mail Date 12/27/04
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application (PTO-152)
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____.
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____.


George Manuel
Primary Examiner
Art Unit: 3762

DETAILED ACTION

Allowable Subject Matter

Claims 1-24 are allowed.

The following is an examiner's statement of reasons for allowance:

Determining the extent to which different myocardial regions are hypertrophied based on sensed amplitudes of an electrogram as claimed is not taught nor suggested by the prior art of record.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.


Ben-Haim '898 and '094 (both of record) teach in an HCM diseased heart, the septum is enlarged and that pacing in an HCM diseased heart causes the mass near the pacing location to be reduced. In addition, a few weeks after an initial pacing regime is used, the heart is imaged to determine morphological changes in the heart. These teachings appear to indicate it would be beneficial to determine the extent to which different myocardial regions are hypertrophied, but lack a teaching or suggestion to incorporate measuring amplitudes of an electrogram in a cardiac rhythm management device. Likewise, Kramer et al '988 (of record) teach stimulating a heart to effect myocardial remodeling; however, there lacks a teaching or suggestion to measure an electrogram or sense during a depolarization to determine the extent to which different myocardial regions are hypertrophied.

Application/Control Number: 10/071,875

Page 3

Art Unit: 3762

Any inquiry concerning this communication or earlier communications from the examiner should be directed to George Manuel whose telephone number is (703) 308-2118.


George Manuel
Primary Examiner
Art Unit: 3762